





Conformity to Type based on Internal Production Control plus supervised product checks at random intervals

This is to certify that: Langshi Garments & Weaving

Co.,Ltd. Jinjiang

Sanou Industrial Zone, Yinglin

Jinjiang Fujian 362200 China

Holds Certificate Number: CE 728635

In respect of:

For the manufacture of particulate respirators to technical specification to satisfy Annex II of the PPE Regulation (EU) 2016/425.

on the basis that BSI carried out the supervised production checks at random intervals under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex VII (Module C2)

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Previous Notified Body: BSI 0086

First Issued: 2020-07-23 Latest Issue: 2020-07-23 Drs. Dave Hagenaars, Managing Director

Effective Date: 2020-07-23 Expiry Date: 2021-07-23

Page: 1 of 3



...making excellence a habit.™

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.

To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.

Conformity to Type based on Internal Production Control plus supervised product checks at random intervals

No. CE 728635

Product manufactured by:

Langshi Garments & Weaving Co.,Ltd. Jinjiang Sanou Industrial Zone, Yinglin Jinjiang **Fujian** 362200 China

Product details

The respiratory protective device covered by the scope of this Module C2 Certificate and the Technical Specification to which the product is manufactured are as follows:

Product type: Particulate filtering half masks for use by Healthcare professionals.

Model and classifications: MSK2 FFP2

Technical Specification: Technical specification to satisfy Annex II of the PPE Regulation (EU) 2016/425.

BSI's PPE for Healthcare Professionals 2020/403 - RPE Technical Specification.

First Issued: 2020-07-23 Effective Date: 2020-07-23 Latest Issue: 2020-07-23 Expiry Date: 2021-07-23

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request. To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.

Conformity to Type based on Internal Production Control plus supervised product checks at random intervals

No. CE 728635

Certificate Administration Details:

Certificate Amendment Record and BSI internal Review relating to this Certificate

Issue date	Comments	BSI Review No.
July 2020	First issue.	2797:20:3204352

Certificate validity

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspects of the overall quality system utilized in the manufacture of the products, failure to do so could invalidate the Certificate in respect of product manufactured after the introduction of such changes.

First Issued: 2020-07-23 Effective Date: 2020-07-23 Latest Issue: 2020-07-23 Expiry Date: 2021-07-23

Page: 3 of 3

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request. To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.

Section 9 **EU Declaration of Conformity**

EU DECLARATION OF CONFORMITY

- 1. PPE (product, type, batch or serial number): MODEL MSK2,product code:2020PM02
- 2. Name and address of the manufacturer and, where applicable, his authorised representative: Langshi Garments & Weaving Co.,Ltd.Jinjiang/Sanou Industrial Zone,Yinglin,Jinjiang,Fujian,China
- 3. This declaration of conformity is issued under the sole responsibility of the manufacturer: Langshi Garments & Weaving Co., Ltd. Jinjiang
- 4. Object of the declaration (identification of PPE allowing traceability; where necessary for the identification of the PPE, a colour image of sufficient clarity may be included): MODEL MSK2 Particulate Respirator
- 5. The object of the declaration described in point 4 is in conformity with the relevant Union harmonisation legislation: PPE Regulation 2016/425
- 6. References to the relevant harmonised standards used, including the date of the standard, or references to the other technical specifications, including the date of the specification, in relation to which conformity is declared: EN149:2001+A1:2009
- 7. Where applicable, the notified body BSI Group NB 0086 performed the EU type-examination (Module B) and issued the EU type-examination certificate CE 600016 EU Type Examination Certificate.
- 8. Where applicable, the PPE is subject to the conformity assessment procedure -conformity to type based on internal production control plus supervised product checks at random intervals (Module C2) under surveillance of the notified body BSI Group NB number:2797.
- 9. Additional information:

Signed for and on behalf of: Langshi Garments & Weaving Co., Ltd. Jinjiang

3. (place and date of issue): Sanou Industrial Zone, Yinglin, Jinjiang, Fujian, China

(name, function) Sam Ouyang, Manager

(signature):



Test Report 3248298.

Langshi Garments & Weaving Co. Ltd.



Introduction.

This report has been prepared by Ben Hobbs and relates to the activity detailed below:

Job/Registration Details		Client Details
Job number: Job type: Start Date: Test type: Sample ID: Registration: Scheme:	3248298 Testing Samples Submitted 06/06/2020 Type 10190967 CE 728634 Negative pressure RPE	Langshi Garments & Weaving Co., Ltd. Jinjiang Sanou Industrial Zone Yinglin Jinjiang Fujian 362200 China
Protocol: Scheme Manager:	PP123 Nathan Shipley	

The report has been approved for issue by T Wicksey - Senior Test Engineer

Approved For Issue	
ZDK	
	Issue Date: 30 June 2020

Objectives.

This is an independent test evaluation to only certain clauses or sub-clauses of the agreed specification in accordance with the following test programme:

BSI COVID-19 filtering face piece technical specification, for COVID-19 masks for use by healthcare workers

Product Scope.

COVID-19 masks for use by healthcare workers

Report Summary.

The samples were received on 3 June 2020 and the testing was started on 6 June 2020.

The samples submitted complied with the requirements of the test work conducted.



Test Samples.

Sample ID	ER Number	Description
1 to 19	10190967	Model: KN95 mask

Description of Test Samples.

Samon	OCCE	mtion
עיוווייע	(=>(=)	ption

Model: KN95. COVID-19 masks for use by healthcare workers





Glossary of Terms.

Pass: Complies. Tested by BSI engineers at BSI laboratories

Pass 1: Complies. Witnessed by BSI engineers in manufacturers laboratory.

Pass 2: Complies. Tests carried out by third party lab; results accepted by BSI.

Pass*: Report resulted in uncertainty and states that Compliance is more probable than non-compliance.

Fail: Non-compliance. Product does not meet the requirements of this clause.

Fail*: Report resulted in uncertainty and states that Non-compliance is more probable than compliance.

N/T: Not Tested N/A: Not Applicable AR: As Received

TC: Temperature Conditioned

SW: Simulated Wear FT: Flow Tested

MS: Mechanical strength

MMDF: Manufactures Minimum Design Flow MMDC: Manufactures Minimum Design Condition

Conditions of Issue.

This Test Report is issued subject to the conditions stated in current issue of 'BSI Terms of Service'. The results contained herein apply only to the particular sample(s) tested and to the specific tests carried out, as detailed in this Test Report. The issuing of this Test Report does not indicate any measure of Approval, Certification, Supervision, Control or Surveillance by BSI of any product. No extract, abridgement or abstraction from a Test Report may be published or used to advertise a product without the written consent of BSI, who reserve the absolute right to agree or reject all or any of the details of any items or publicity for which consent may be sought.

Should you wish to speak with BSI in relation to this report, please contact Customer Services on +44 (0)8450 80 9000.

BSI Kitemark House Maylands Avenue Hemel Hempstead Hertfordshire HP2 4SQ



Opinions and Interpretations expressed herein are outside the scope of our UKAS accreditation.

Unless otherwise stated, any results not obtained from testing in a BSI laboratory are outside the scope of our UKAS accreditation.



Test Results.

Testing in accordance with BSI COVID-19 filtering face piece technical specification

BS EN 149:2001 +A1:2009 Technical testing specification for COVID-19 masks for use by healthcare workers

CLAUSE	REQUIREMENTS	ASSESSMENT
7.7	Practical performance	
	The particle filtering half mask shall undergo practical performance tests under realistic conditions. These general tests serve the purpose of checking the equipment for imperfections that cannot be determined by the tests described elsewhere in this standard.	
	Where practical performance tests show the apparatus has imperfections related to wearer's acceptance, the test house shall provide full details of those parts of the practical performance tests which revealed these imperfections.	
	Test in accordance with clause 8.4 of the standard.	Pass
	Testing in accordance with BSI COVID-19 filtering face piece technical specification, for masks for use by healthcare workers During the tests the particle filtering half mask shall be subjectively assessed by the wearer and after the test, comments on the following shall be recorded:	

Table A: Practical performance

Toot			Co	omments		
Test candidate	Sample	Head harness comfort	Security of fastenings	Field of vision	Any other comments	Assessment
RF1	1 AR	OK	OK	OK	None	Pass
JS3	2 AR	OK	OK	OK	None	Pass

a) head harness comfort; b) security of fastenings; c) field of vision; d) any other

7.9 Leakage

7.9.1 Total inward leakage

The laboratory tests shall indicate that the particle filtering half mask can be used by the wearer to protect with high probability against the potential hazard to be expected.

The total inward leakage consists of three components: face seal leakage, exhalation valve leakage (if exhalation valve fitted) and filter penetration.

Test in accordance with clause 8.5 of the standard.

comments reported by the wearer on request.

Pass

Testing in accordance with BSI COVID-19 filtering face piece technical specification, for masks for use by healthcare workers

5 test subjects, masks tested 'As received'. All individual exercise results tests shall be not greater than 11 % (for FFP2) and, in addition, all arithmetic means for the total inward leakage shall be not greater than 8 % (for FFP2).

Table B: Clause 7.9.1 - Total inward leakage

Table B. Cla	usc 7.5.1	i otai iiiwa	ra icanage						
		Samuel Pre test	Inward Leakage (%)						
Test	Cample		Α	В	С	D	E		
candidate	Sample	condition	Walking	Walking with head side to side	Walking with head up & down	Walking and talking	Walking	Average	Assessment
JB1	3	AR	5.0281	7.3048	10.0191	7.0132	7.0601	7.2850	Pass
AH1	4	AR	6.5032	6.5831	7.3376	4.3008	7. 444 7	6.4339	Pass
JS3	5	AR	3.1955	3.3215	5.1665	2.7727	3.7286	3.6369	Pass
BH1	6	AR	7.3943	7.9196	7.0189	5.4025	9.4578	7.4386	Pass
CB1	7	AR	3.7106	3.8195	4.1821	4.2392	5.0839	4.2071	Pass



Test Results. (Continued)

CLAUSE	REQUIREMENTS	ASSESSMENT
--------	--------------	------------

7.9.2 Penetration of filter material

Testing in accordance with BSI COVID-19 filtering face piece technical specification, for masks for use by healthcare workers

3 test samples masks tested 'As received', for NaCl (Sodium Chloride) and PO (Paraffin oil), 3 min test. Testing shall be done in accordance with 8.11. 6% limit for both PO and NaCl

Pass

Table C: Clause 8.11 - Sodium Chloride penetration test

Table of clause 0.11 Socialit enionae penetration test								
Sample	Pre-test	Flow through filter (I/min)	Penetra	ation (%)				
number	condition	Flow through filter (I/min)	Limit	Actual				
8	AR			1.1782				
9	AR	95	< 6	1.2865				
10	AR			1.2686				

Table D: Clause 8.11 - Paraffin oil penetration test

Sample	Pre-test	Flow through filter (I/min)	Penetra	ation (%)				
number	condition	Flow tillough filter (i/illill)	Limit	Actual				
11	AR			2.1190				
12	AR	95	< 6	1.8915				
13	AR			2.5045				

7.12 Carbon dioxide content of inhalation air

The carbon dioxide content of the inhalation air (dead space) shall not exceed an average of 1.0% (by volume).

Test in accordance with clause 8.7 of the standard.

Pass

Table E: Clause 8.7 - Carbon Dioxide content of the inhalation air

Camania	Due test sandition	Dead space CO ₂ (%)		
Sample	Pre-test condition	Limit	Measured	
14	AR		0.49	
15	AR	< 1.0	0.54	
16	AR		0.44	



Test Results. (Continued)

CLAUSE	REQUIREMENTS	ASSESSMENT
--------	--------------	------------

7.16 Breathing resistance

Testing in accordance with BSI COVID-19 filtering face piece technical specification, for masks for use by healthcare workers

3 test samples masks tested 'As received'. Test in accordance with clause 8.9 of the standard.

Pass

The breathing resistances shall meet the requirements of FFP2; 30l/min – 0.7mbar (inhale), 95l/min – 2.4mbar (inhale), 160l/min – 3.0mbar (exhale)

Table F: Clause 8.9 – Breathing resistance. Inhalation resistance at a continuous flow

Cample	Pre-test	Continuous flow (I/min)	Inhalation resistance (mbar)		
Sample	condition		Limit	Measured	
17	AR		< 0.7	0.40	
18	AR	30		0.41	
19	AR			0.35	
17	AR			1.23	
18	AR	95	< 2.4	1.32	
19	AR			1.19	

Table G: Clause 8.9 – Breathing resistance. Exhalation resistance at a continuous flow, measured in five orientations with the worst case reported

Cample	Pre-test condition	Continuous flow (I/min)	Exhalation resistance (mbar)		
Sample			Limit	Measured	
17	AR		< 3.0	2.10	
18	AR	160		2.23	
19	AR			2.03	



Appendix A. – Test Panel Data

Test Candidate	Facial Dimensions (mm)					Sex
	Length of face	Width of face	Face depth	Width of mouth	Head Circumference	Sex
JB1	114	144	108	59	574	Male
JS3	126	134	124	49	600	Male
BH1	114	139	120	50	570	Male
CB1	117	147	130	57	566	Male
AH1	108	124	130	46	570	Male
RF1	104	122	121	55	549	Male

Note: All candidates were clean shaven


Product photographs.



Front view



Side View



End of Report